

# DHEA-S ELISA TEST SYSTEM

## INTENDED USE

DHEA-S ELISA Test System is used for the quantitative measurement of DHEA-S in human serum and plasma.

## SUMMARY AND EXPLANATION

DHEA-S is a more specific product of the adrenals and measurements of this steroid are widely intended in clinical practice. The clinical importance of plasma assays of DHEA-S is associated with the diagnosis of adrenal hyperplasia and differential diagnosis of hirsutism.

## PRINCIPLE OF THE TEST

DHEA-S kit is based on the principle of competitive binding between DHEA-S in the test specimen and HRP conjugated DHEA-S for a constant amount of anti-DHEA-S antibody. In the first incubation, Goat-anti-Rabbit-IgG coated wells are incubated with 10µl of DHEA-S standards, patient samples, 50µl DHEA-S Enzyme (HRP) reagent and 50µl anti-DHEA-S Antibody reagent, at room temperature, for 60 minutes. During the incubation, HRP labeled DHEA-S competes with the endogenous DHEA-S in the standard and sample, for a fixed number of binding sites of the DHEA-S antibody, while simultaneously the Anti DHEA-S antibody binds to the immobilized secondary antibody. Thus, the amount of DHEA-S HRP conjugate immunologically bound to the well progressively decreases as the concentration of DHEA-S in the specimen increases. Unbound DHEA-S HRP conjugate is then removed and the wells washed. Next, TMB Reagent is added and incubated at room temperature for 30 minutes, resulting in the development of blue color. The color development is stopped with the addition of stop solution, and the absorbance is spectrophotometrically measured at 450nm. A standard curve is prepared relating color intensity to the concentration of DHEA-S.

## MATERIALS AND COMPONENTS PROVIDED

• Microwells coated with Goat anti-Rabbit IgG	12x8x1
• Standard set, 7 vials (ready to use)	0.25 ml
• DHEA-S Enzyme Reagent, 1 bottle (ready to use)	6 ml
• Anti-DHEA-S Antibody Reagent, 1 bottle (ready to use)	6 ml
• TMB Substrate (ready to use)	12 ml
• Stop Solution (ready to use)	12 ml
• Wash solution, 1 bottle (20X)	25 ml
• Control Level 1	0.25 ml
• Control Level 2	0.25 ml

## MATERIALS REQUIRED BUT NOT PROVIDED

- Precision pipettes
- Disposable pipette tips
- ELISA reader capable of reading absorbance at 450 nm
- Flat-head Vortex mixer
- Plate shaker
- Graph paper

## STORAGE CONDITIONS

- Store the kit at 2-8°C.
- Keep microwells sealed in a dry bag with desiccants.
- The reagents are stable until expiration of the kit.
- Do not expose test reagents to heat, sun or strong light.

## PRECAUTIONS

1. Potential biohazardous materials: The calibrator and controls contain human source components, which have been tested and found non-reactive for hepatitis B surface antigen as well as HIV antibody with FDA licensed reagents. However, there is no test method that can offer complete assurance that HIV, Hepatitis B virus or other infectious agents are absent. These reagents should be handled at the Biosafety Level 2, as recommended in the Centers for Disease Control/National Institutes of Health manual, "Biosafety in Microbiological and Biomedical Laboratories" 1984.
2. This test kit is USA FDA exempt product.
3. Do not pipette by mouth. Do not smoke, eat, or drink in the areas in which specimens or kit reagents are handled.
4. The components in this kit are intended for use as an integral unit. The components of different lots should not be mixed.
5. It is recommended that standards, control and serum samples be run in duplicate.
6. Optimal results will be obtained by strict adherence to this protocol. Accurate and precise pipetting, as well as following the exact time and temperature requirements prescribed are essential. Any deviation from this may yield invalid data.

## SPECIMEN COLLECTION

1. Collect blood specimens and separate the serum immediately.
2. Specimens may be stored refrigerated at (2-8°C) for 5 days. If storage time exceeds 5 days, store frozen at (-20°C) for up to one month.
3. Avoid multiple freeze-thaw cycles.
4. Prior to assay, frozen sera should be completely thawed and mixed well.
5. Do not use grossly lipemic specimens.

## REAGENT PREPARATION

20X Wash Buffer Concentrate: Prepare 1X wash buffer by adding the contents of the bottle to 475 ml of distilled or deionized water. Store at room temperature (18-24°C).

## TEST PROCEDURE

All reagents and specimens must be allowed to come to room temperature before use.

1. Secure the desired number of coated strips in the holder.
2. Dispense 10 µl of each standards, controls and sample with new

disposable tips into appropriate wells.

3. Dispense 50 µl of DHEA-S Enzyme-Reagent into each well.
4. Dispense 50 µl of Anti-DHEA-S Antibody-Reagent into each well.
5. Thoroughly mix the plate for 10 seconds. It is important to have complete mixing in this step.
6. Incubate for 60 minutes at room temperature.
7. Briskly shake out the contents of the wells.
8. Rinse the wells 3 times with diluted Wash Solution (350 µl per well). Strike the wells sharply on absorbent paper to remove residual droplets.
9. Add 100 µl of TMB Substrate into each well.
10. Incubate for 30 minutes at room temperature.
11. Stop the enzymatic reaction by adding 50 µl of Stop Solution into each well.
12. Read absorbance on ELISA Reader at 450 nm within 10 minutes of adding the Stop Solution.

## CALCULATION OF RESULTS

The standard curve is constructed as follows:

1. Check DHEA-S standard value on each standard vial. This value might vary from lot to lot. Make sure you check the value on every kit. See example of the standard attached.
2. To construct the standard curve, plot the absorbance for DHEA-S standards (vertical axis) versus standard concentrations (horizontal axis) on a linear graph paper. Draw the best curve through the points.
3. Read the absorbance for controls and each unknown sample from the curve. Record the value for each control or unknown sample.

Example of a Standard Curve

	Conc. µg/mL	OD 450 nm
Std 1	0	2.02
Std 2	0.1	1.34
Std 3	0.5	0.95
Std 4	1	0.70
Std 5	2.5	0.40
Std 6	5	0.24
Std 7	10	0.13

## TEST LIMITATION

1. The test results obtained using this kit serve only as an aid to diagnosis and should be interpreted in relation to the patient's history, physical findings and other diagnostic procedures.
2. Do not use sodium azide as preservative. Sodium azide inhibits HRP enzyme activities.

## EXPECTED VALUES

It is recommended that each laboratory establish its own normal ranges based on a representative sampling of the local population. The following values may be used as initial guideline ranges only:

Classification	Normal Range
Male	1.0 - 4.2 µg/ml
Female	
Premenopausal	0.8 - 3.9 µg/ml
Term Pregnancy	0.2 - 1.2 µg/ml
Postmenopausal	0.1 - 0.6 µg/ml
Newborn (both sexes)	1.7 - 3.6 µg/ml

Conversion factor: 1 µg/ml = 2.6 µmol/L

## PERFORMANCE CHARACTERISTICS

### 1. Correlation with a Reference ELISA kit:

A total of 60 sera were tested by this ELISA and a reference ELISA kit. Results were as follows:

Correlation	Slope	Intercept
0.99	1.05	0.026

### 2. Precision

#### Intra-Assay

Serum	No. of Replicates	Mean µg/mL	Standard Deviation	Coefficient of Variation (%)
1	16	0.82	0.033	4.0
2	16	1.61	0.061	3.8
3	16	4.75	0.169	3.5

#### Inter-Assay

Serum	No. of Replicates	Mean µg/mL	Standard Deviation	Coefficient of Variation (%)
1	16	0.82	0.035	4.23
2	16	1.59	0.053	3.35
3	16	4.68	0.177	3.77

### 3. Sensitivity

The sensitivity was determined by calculating the mean plus 2SD of the standard zero point tested 20 times in the same run.

Serum	No. of Replicates	Mean (µg/ml)	Standard Deviation	Mean + 2SD (Sensitivity)
Zero Standard	20	0.007	0.0165	0.0236

### 4. Specificity

The % cross reactivity of the DHEA-S test, to other related compounds was evaluated by spiking the compounds to serum matrix at massive concentrations. The cross reactivity was computed by deriving a ratio between dose of interfering compound to dose of DHEA-S needed to displace the same amount of labeled analog.

Compound	% Cross reactivity
DHEA-S	100%
E1, 10ng/ml	ND
E2, 10ng/ml	ND
E3, 400ng/ml	ND
Progesterone, 400ng/ml	ND
Cortisol, 8µg/ml	ND
Testosterone, 200ng/ml	ND
Androstenedione, 1µg/ml	ND
DHEA, 300ng/ml	ND

## REFERENCES

1. Tietz, N. W., Textbook of Clinical Chemistry, Saunders, 1968

## PRESENTACIÓN:

CONT. 96 TEST CODIGO: RSET072-2